


Specifications for ThyroChek®



120-A

Box of 20 tests, pipettes, buffer, inserts



120-B

Self-sealing bag of 20 tests, pipettes, buffer, inserts



120-C

Positive and negative controls

- **Shipper (Case):** 20 Bags or 12 boxes 17.7 by 43.2 by 28 cm or 12 by 18 by 20"
Storage: Room temperature 15-30°C/60-86°F. Do not freeze.
Expiry dating: minimum 18 months from date shipped.

Shipping: FOB San Diego

Controls: Available as Product Number 120-C.

Proficiency Testing: registered with major services.

Specificity: 96.95%

Sensitivity: 98.5%

Precision: 97.7%

CLIA waived: approved October 4, 2006

Moderate Complexity: 510 K 990658

UPC code: application made

CE mark: application made

US Patent: 5821073

US manufacture: ISO 13485:1996

CPT Code: 84443QW

Reimbursement: \$27.47 in most States

TO PLACE ORDER

- **FAX: 801-720-7568**
- **EMAIL: info@cliawaived.com**
- **TEL: 858-481-5031**

Payments should be made out and sent to our mailing address:

CLIAwaived, Inc.
4332 Corte de la Fonda
San Diego, CA 92130